



CANON INC.

9-1, Imaikami-cho, Nakahara-ku, Kawasaki, Kanagawa 211-8501, Japan  
Phone +81-3-3758-2111

Ref. No. M1TJ-21R030 rev.02

## UK DECLARATION OF CONFORMITY

Manufacturer: CANON INC.

9-1, Imaikami-cho, Nakahara-ku, Kawasaki, Kanagawa 211-8501, Japan

declare under our sole responsibility that the product(s)

Product Name: DIGITAL RADIOGRAPHY

model: CXDI-710C Wireless CXDI-810C Wireless CXDI-410C Wireless

CXDI-710C Wireless (WM5A11), CXDI-810C Wireless (WM5A12), CXDI-410C Wireless (WM5A13)

is(are) in conformity with the essential requirements of

S.I. 2017 No. 1206 & S.I. 2012 No. 3032

by applying the following standards

UK Regulation(s)	Reference of standard(s) and amendment(s)
S.I. 2017 No. 1206 (Radio Equipment Regulations)	EN 300 328 V2.2.2
	EN 301 893 V2.1.1
	EN 50566: 2017
	EN 62368-1:2014 + A11:2017
	EN 301 489-17 V3.2.4
	EN 301 489-1 V2.2.3
S.I. 2012 No. 3032 (RoHS)	EN IEC 63000:2018

Note:

1. The quality system covering the production is implemented according to ISO EN13485 or monitored based on appropriate measures.
2. The internal design control system is implemented according to ISO 14000-series and monitored based on appropriate measures.

Place : Kawasaki, Japan

Date: 20th February, 2023

Hideki Nonaka

General Manager

Medical Equipment Development Promotion Div.

Medical Components Business Unit

CANON INC.